laboratory. In the latter case, the blood shall be placed in temporary storage having sufficient refrigeration capacity to cool the blood continuously toward a range between 1 and 6 °C until it arrives at the processing laboratory, where it shall be stored at a temperature between 1 and 6 °C. Blood from which Platelets is to be prepared shall be held in an environment maintained at a temperature range 20 to 24 °C until the platelets are separated. The red blood cells shall be placed in storage at a temperature between 1 and 6 °C immediately after the platelets are separated.

[38 FR 32089, Nov. 20, 1973, as amended at 42 FR 59878, Nov. 22, 1977; 43 FR 34460, Aug. 4, 1978; 49 FR 23834, June 8, 1984; 50 FR 4138, Jan. 29, 1985; 55 FR 11013, Mar. 26, 1990; 64 FR 45372, Aug. 19, 1999; 66 FR 1836, Jan. 10, 2001; 66 FR 40889, Aug. 6, 2001]

§ 640.5 Testing the blood.

All laboratory tests shall be made on a specimen of blood taken from the donor at the time of collecting the unit of blood, and these tests shall include the following:

(a) *Serological test for syphilis.* Whole Blood shall be negative to a serological test for syphilis.

(b) *Determination of blood group.* Each container of Whole Blood shall be classified as to ABO blood group. At least two blood group tests shall be made and the unit shall not be issued until grouping tests by different methods or with different lots of antiserums are in agreement. Only those Anti-A and Anti-B Blood Grouping Reagents licensed under, or that otherwise meet the requirements of, the regulations of this subchapter shall be used, and the technique used shall be that for which the serum is specifically designed to be effective.

(c) *Determination of the Rh factors.* Each container of Whole Blood shall be classified as to Rh type on the basis of tests done on the sample. The label shall indicate the extent of typing and the results of all tests performed. If the test, using Anti-D Blood Grouping Reagent, is positive, the container may be labeled ''Rh Positive.'' If the test is negative, the results shall be confirmed by further testing which shall include tests for the ''weak D (formerly D^u).'' Blood may be labeled ''Rh Negative'' if further testing is negative. Units testing positive after additional more specific testing shall be labeled as ''Rh Positive.'' Only Anti-Rh Blood Grouping Reagents licensed under, or that otherwise meet the requirements of, this subchapter shall be used, and the technique used shall be that for which the reagent is specifically designed to be effective.

(d) *Sterility test.* Whole Blood intended for transfusion shall not be tested for sterility by a method that entails entering the final container before the blood is used for transfusion.

(e) *Inspection.* Whole Blood shall be inspected visually during storage and immediately prior to issue. If the color or physical appearance is abnormal or there is any indication or suspicion of microbial contamination the unit of Whole Blood shall not be issued for transfusion.

(f) *Test for communicable disease agents.* Whole Blood shall be tested for evidence of infection due to communicable disease agents as required under § 610.40 of this chapter.

[38 FR 32089, Nov. 20, 1973, as amended at 50 FR 4138, Jan. 29, 1985; 53 FR 117, Jan. 5, 1988; 53 FR 12764, Apr. 19, 1988; 64 FR 45372, Aug. 19, 1999; 66 FR 1836, Jan. 10, 2001; 66 FR 31165, June 11, 2001; 66 FR 40889, Aug. 6, 2001]

§ 640.6 Modifications of Whole Blood.

Upon approval by the Director, Center for Biologics Evaluation and Research, of a supplement to the biologics license application for Whole Blood a manufacturer may prepare Whole Blood from which the antihemophilic factor has been removed, provided the Whole Blood meets the applicable requirements of this subchapter and the following conditions are met:

(a) The antihemophilic factor shall be removed in accordance with paragraphs (a), (b), and (c) of § 640.52.

(b) Although the closed system between the red blood cells and plasma shall be maintained, the red blood cells shall be maintained between 1 and 6° C at all times, including that time when

the plasma is being frozen for removal of the antihemophilic factor.

[38 FR 32089, Nov. 20, 1973, as amended at 49 FR 23834, June 8, 1984; 50 FR 4138, Jan. 29, 1985; 55 FR 11013, Mar. 26, 1990; 59 FR 49351, Sept. 28, 1994; 64 FR 45372, Aug. 19, 1999; 64 FR 56453, Oct. 20, 1999]

Subpart B—Red Blood Cells

§ 640.10 Red Blood Cells.

The proper name of this product shall be Red Blood Cells. The product is defined as red blood cells remaining after separating plasma from human blood.

[38 FR 32089, Nov. 20, 1973, as amended at 50 FR 4138, Jan. 29, 1985]

§ 640.11 General requirements.

(a) *Storage.* Immediately after processing, the Red Blood Cells shall be placed in storage and maintained at a temperature between 1 and 6 °C.

(b) *Inspection.* The product shall be inspected immediately after separation of the plasma, periodically during storage, and at the time of issue. The product shall not be issued if there is any abnormality in color or physical appearance or if there is any indication of microbial contamination.

[38 FR 32089, Nov. 20, 1973, as amended at 41 FR 18292, May 3, 1976; 42 FR 59878, Nov. 11, 1977; 50 FR 4139, Jan. 29, 1985]

§ 640.12 Suitability of donor.

The source blood for Red Blood Cells shall be obtained from a donor who meets the criteria for donor suitability prescribed in § 640.3.

[38 FR 32089, Nov. 20, 1973, as amended at 50 FR 4139, Jan. 29, 1985]

§ 640.13 Collection of the blood.

(a) The source blood shall be collected as prescribed in § 640.4.

(b) Source blood may also be derived from Whole Blood manufactured in accordance with applicable provisions of this subchapter.

[38 FR 32089, Nov. 20, 1973, as amended at 50 FR 4139, Jan. 29, 1985; 64 FR 45372, Aug. 19, 1999]

§ 640.14 Testing the blood.

Blood from which Red Blood Cells are prepared shall be tested as prescribed in § 610.40 of this chapter and § 640.5 (a), (b), and (c).

[53 FR 117, Jan. 5, 1988, as amended at 66 FR 31165, June 11, 2001]

§ 640.15 Segments for testing.

Segments collected in integral tubing shall meet the following standards:

(a) One or more segments shall be provided with each unit of Whole Blood or Red Blood Cells when issued or reissued.

(b) Before they are filled, all segments shall be marked or identified so as to relate them to the donor of that unit of red cells.

(c) All segments accompanying a unit of Red Blood Cells shall be filled at the time the blood is collected or at the time the final product is prepared.

[66 FR 40890, Aug. 6, 2001]

§ 640.16 Processing.

(a) *Separation.* Within the timeframe specified in the directions for use for the blood collecting, processing, and storage system used, Red Blood Cells may be prepared either by centrifugation, done in a manner that will not tend to increase the temperature of the blood, or by normal undisturbed sedimentation. A portion of the plasma sufficient to insure optimal cell preservation shall be left with the red cells except when a cryoprotective substance or additive solution is added for prolonged storage.

(b) *Sterile system.* All surfaces that come in contact with the red cells shall be sterile and pyrogen-free.

(c) *Final containers.* Final containers used for Red Blood Cells shall be the original blood containers unless the method of processing requires a different container. The final container shall meet the requirements for blood containers prescribed in § 640.2(c). At the time of filing, if a different container is used, it shall be marked or identified by number or other symbol so as to relate it to the donor of that unit of red cells.

[38 FR 32089, Nov. 20, 1973, as amended at 43 FR 34460, Aug. 4, 1978; 50 FR 4139, Jan. 29, 1985; 64 FR 45372, Aug. 19, 1999; 66 FR 1836, Jan. 10, 2001; 66 FR 40890, Aug. 6, 2001]